

# REF. 11584 LOGÍSTICA

#### Cajas de 48 Paquetes

Paquetes de 20 Unidades (Embolsadas Individualmente) Caja cartón 50 x 30 x 33 cm

PESO CAJA 6 KILOS



**NB 2163** 

### **EU TYPE EXAMINATION CERTIFICATE**

Certificate No: 2163-PPE-882

Respiratory protective devices, filtering half masks to protect against particles manufactured by

Zhejiang Yiju Medical Instrument Co., Ltd.

Songqiao Industrial Zone, Wanquan Town, Pingyang County, Wenzhou, Zhejiang, China are tested and evaluated according to

EN 149:2001 + A1:2009 Respiratory Protective Devices -Filtering Half Masks to Protect Against Particles -Requirements, Testing, Marking

Based on the type examination conducted with the evaluation of test reports, technical file according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 5, it is approved that the product meets the requirements of the regulation.

#### **Product Definition**

Brand Name: ZJYIJU Model: YJ20-01 Filtering half mask Classification: FFP2 NR

Here by the manufacturer is allowed to use notified body number (2163) and can fix CE mark, as shown below, on the Category III product models given above, with;

- Issuing an appropriate EU Declaration of Conformity according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 9.
- Ongoing successful performance in fulfilment of the requirements set out in Personal Protective Equipment Regulation (EU) 2016/425 and harmonised standards, ensured by assessments based on Annex 7 (Module C2) or Annex 8 (Module D) of the regulation no later than 1 year from the beginning of serial production

This certificate is initially issued on 29/06/2020 and will be valid for 5 years, if there is no change in the relevant harmonised standard affecting the essential health and safety requirements.



Suat KAÇMAZ
UNIVERSAL CERTIFICATION
Director





**NB 2163** 

## **CERTIFICATE OF CONFORMANCE**

Certificate No: 2163-PPE-882/01

Respiratory protective devices, filtering half masks to protect against particles manufactured by

Zhejiang Yiju Medical Instrument Co., Ltd.

Songqiao Industrial Zone, Wanquan Town, Pingyang County, Wenzhou, Zhejiang, China

Continues to fulfil the requirements of

EN 149:2001 + A1:2009 Respiratory Protective Devices -Filtering Half Masks to Protect Against Particles -Requirements, Testing, Marking

Based on the evaluation of test reports and internal quality control audit reports according to EN 149+A1:2009 and Personal Protective Equipment Regulation (EU) 2016/425 Annex VII (Module C2). This certificate implies that the manufactured products show below are in conformance with the approved EU Type Examination model and meets the requirements of the regulation.

**Product Definition** 

Model	Class	EU Type Examination Certificate		
		Serial No	Date	Issuing NB No
ZJYIJU / YJ20-01	FFP2 NR	2163-PPE-882	29.06.2020	2163

Here by the manufacturer is allowed to use notified body number (2163) and can fix CE mark, as shown below, on the Category III product models given above, with;

- Issuing an appropriate EU Declaration of Conformity according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 9.
- Taking all measures necessary so that the manufacturing process and its monitoring ensure the homogeneity of production and conformity of the manufactured PPE with the type described in the EU type examination certificate.

This certificate is issued on 30/07/2020 and will be valid for one year, until 29/07/2021 if the manufacturer makes no major change in the product designs and manufacturing processes affecting the product performance on the essential health and safety requirement.

**CE** 2163

Suat KAÇMAZ
UNIVERSAL CERTIFICATION
Director